

Bionic Products Pty Ltd
Unit 11/15 John Duncan Court
Varsity Lakes
QLD 4227
Australia

Page 1 of 1

Certificate of Test

Electrical Safety
For
Pendant Ioniser

Date of Issue: 14th December 2015
SGS Reference: ELS209406/3/C/MG/15
Details of Sample: Model: Elanra Air
Rated: IP22
Date of Receipt: 18th September 2015
Date of Test: 8th December 2015
Test Specification: BS EN 60529 + *A2:2013 Degrees of protection provided by enclosures (IP code)
Conclusion: The samples tested were found to comply with the Test Specification.

Test Engineer:



M. W. Garfitt

Counter Signatory:



D. Carroll

All enquires relating to this certificate must be directed to the Test Engineer



This document is issued by the Company subject to its General Conditions of Service available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 28 days only.

Tests marked * are not UKAS accredited