

Bionic Products PTY Ltd.
Unit 11/15 John Duncan Court,
Varsity Lakes, Queensland.
QLD 4227
Australia

Page 1 of 1

Certificate of Test

Electromagnetic Compatibility
for

Elanra Air Pendant Therapeutic Ioniser

Date of Issue: 2nd February 2016

SGS Reference: EMC209406/2

Details of Sample: Model No.: C1000110
Unique Identifier: 000002

Date of Receipt: 18th September 2015

Date of Test: 8th October 2015 To 26th January 2016

Test Required: Test for compliance with: -
EN 60601-1-2: 2007

Modifications: Sample 1 – Original Sample. All other tests were performed with this sample.

Sample 2 – The mains power adaptor model PSAC05R-050 was replaced with model PSAC05R-050L6-R by the client in order to comply with the radiated emissions test.

Conclusion: In the configuration tested the sample meets the requirements of the standards detailed above.

Issue Number: 1

This document is issued by the Company subject to its General Conditions of Service available on request or accessible at www.sgs.com/terms_and_conditions.htm and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's Instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. "Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 28 days only."

Authorised Signatory

G. Hann
Automotive Manager

